

# 39<sup>A</sup> GDS Helpsheet



*World Leaders in Computer Controlled Testing  
Systems for Geotechnical Engineers and Geologists*

## Hardware

### Mid Plane Pore Pressure Probe

### Transducer Access Insert Version

---

## 1. Overview

---

The GDS Mid-Plane Pore Pressure Probe (M4P) is designed to provide measurement of pore pressure on the surface of the mid-height of the triaxial test specimen. Shear strain is concentrated in the middle third of the test specimen and so it is logical to measure pore pressure in this region of maximum activity.

## 2. Installation

---

The kit supplied comprises the following components: Druck PDCR 81 pressure transducer, Cutter and Rubber Block, Flanged Grommet, Pair of o-rings and o-ring stretcher, Right Angle Bracket. Their purpose and installation is described below.

1. Druck PDCR 81 pressure transducer. The transducer is provided with a small ceramic tip. This is the sensing end of the transducer and is placed against the surface of the triaxial test specimen. The transducer is shrouded in heat-shrink sleeving which is connected to a bulkhead gland. The bulkhead gland is provided with an o-ring seal to seal against the outside of the triaxial cell. The flying lead is provided with a LEMO connector which is colour-matched to the GDS Digital Transducer Interface (DTI).
2. Drill a hole in the base or top of the cell to take the insert (**if you have an access ring this hole is already in place**). This hole should have a diameter of  $14.287 \pm 0.025$ mm. It may be necessary to use a spacer or additional washers when screwing down the clamping screw.
3. Cutter and Rubber Block. The cutter is provided to cut a small hole in the middle of the rubber sleeve subsequently used to jacket the triaxial test specimen. Lay the rubber sleeve on a flat surface. The Rubber Block is positioned inside the rubber sleeve at the midpoint. Hold the cutting tool vertically above the block with the brass serrated end touching the rubber sleeve. Press down on the brass top of the cutter against the reaction of the rubber block while rotating the outer body of the cutter. Rotate a few times until a hole is cut in the sleeve through to the rubber block. This hole is exactly the right size to take the flanged grommet. Use a small self-adhesive label to cover the hole to make it airtight while you use a suction sleeve stretcher to apply the sleeve to the test specimen. When you have done this simply peel off the label.
4. Flanged Grommet. The grommet is designed to fit exactly through the hole bored in the rubber membrane jacketing the triaxial test specimen. The flange is designed to lie under the membrane against the surface of the test specimen. The grommet pokes through the membrane to receive the transducer tip.

5. Pair of o-rings and o-ring stretcher. Slide the pair of o-rings over the sensor and onto the transducer lead. Slip the o-ring stretcher over the thinner part of the transducer lead between the sensor and the bulkhead gland with the tapered end away from the sensor. Slide the two o-rings onto the stretcher from the tapered end. Push the sensor into the grommet so that the ceramic tip is against the surface of the test specimen. Slide the o-ring stretcher along the cable, over the sensor, and up against the test specimen. Roll the o-rings off so that the grommet grips the sensor and holds it in place against the surface of the test specimen.
6. Right Angle Bracket. Close to the sensor, slip one end of the Right Angle Bracket onto the cable. Then bend the cable gently around into a right angle and slip the cable into the other end of the Right Angle Bracket. This leads the cable nicely away from the wall of the triaxial cell and down to the bulkhead gland.