

# 120 Helpsheet



*World Leaders in Computer Controlled Testing  
Systems for Geotechnical Engineers and Geologists*

Hardware

Power Supply

## GDS Hardware Power requirements

---

Please note, these figures are to be used as a general guide only.

<b>PRODUCT</b>	<b>POWER (WATTS)</b>
ADVANCED CONTROLLER 2Mpa	130
ADVANCED CONTROLLER 32 & 64Mpa	150
STANDARD CONTROLLER 2Mpa	130
HIGH PRESSURE 128Mpa	1000
DATA LOGGER (SERIAL PAD)	50
MULTIPLEXER (MUX)	50
2 Hz MINIDYN	3000
STATIC MINIDYN	130
VIS 250kN	1250
10Hz DYNAMIC	10000 (3 phase)
DTI	50
DPI	50
AIR VALVE	50
Standard PC + Monitor.	250